

Creatin Kinase. NAC. Kinetic UV. Liquid
Quantitative Determination of Creatin Kinase (CK)
 Only for *in vitro* use in clinical laboratory
 Store at 2-8°C

Ref.: NAC-017

Ratio 4:1



CREATIN KINASE



PRINCIPLE OF THE METHOD

Kinetic determination of the creatine kinase based upon IFCC and DGKC recommendations.

Creatine kinase (CK) catalyses the reversible transfer of a phosphate group from phosphocreatine to ADP. This reaction is coupled to those catalysed by hexokinase (HK) and glucose-6-phosphate dehydrogenase (G6P-DH):



The rate of NADPH formation, measured photometrically, is proportional to the catalytic concentration of CK present in the sample^{1,2}.

CLINICAL SIGNIFICANCE

Creatine kinase is a cellular enzyme with wide tissue distribution in the body. Its physiological role is associated with adenosine triphosphate (ATP) generation for contractile or transport systems.

Elevated CK values are observed in diseases of skeletal muscle and after myocardial infarction^{1,5,6}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENTS

R 1	Imidazol, pH 6.7	125 mmol/L
	D-Glucose	25 mmol/L
	N-Acetyl-L-Cysteine	25 mmol/L
	Magnesium acetate	12.5 mmol/L
	NADP	2.52 mmol/L
	EDTA	2.02 mmol/L
R 2	Hexokinase	≥ 800 U/L
	ADP	15.2 mmol/L
	AMP	25 mmol/L
	di-Adenosine-5- pentaphosphate	103 mmol/L
	Glucose-6-phosphate dehydrogenase	≥ 800 U/L
	Creatine phosphate	250 mmol/L

Optional

CK-Nac / CK-MB CONTROL	Lyophilized human serum
------------------------	-------------------------

PREPARATION

Mix 4 volumes of reagent 1 with 1 volume of reagent 2.

Stability: 2 weeks at 2-8°C or 48 hours at room temperature (15-25°C).

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.

Do not use the tablets if appears broken.

Signs of reagent deterioration:

- Presence of particles and turbidity.

- Blank absorbance (A) at 340 nm ≥ 1.60.

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 340 nm.
- Thermostatic bath at 25°C, 30°C ó 37°C (± 0.1°C).
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

Serum free of hemolysis or heparin plasma.

Stability 7 days at 2-8°C, protected from light.

The creatin kinase activity decreases 10% after 1 day at 2-5°C or after 1 hour at 15-25°C.

PROCEDURE

1. Assay conditions:
 Wavelength: 340 nm
 Cuvette: 1 cm light path
 Constant temperature 25°C / 30°C / 37°C
2. Adjust the instrument to zero with distilled water.
3. Pipette into a cuvette:

	25 - 30°C	37°C
WR (mL)	1.0	1.0
Sample (µL)	40	20

3. Mix and incubate 2 minutes.
4. Read initial absorbance (A) of the sample, start the stopwatch and read absorbances at 1 minute intervals thereafter for 3 minutes.
5. Calculate the difference between absorbances and the average absorbance differences per minute (ΔA/min).

CALCULATIONS

25°- 30°C ΔA / min x 4127 = U/L CK

37°C ΔA / min x 8095 = U/L CK

Units: One international unit (IU) is the amount of enzyme that transforms 1 µmol of substrate per minute, in standard conditions. The concentration is expressed in units per litre of sample (U/L).

Temperature conversion factors

To correct results to other temperatures multiply by:

Assay temperature	Conversion factor to		
	25°C	30°C	37°C
25°C	1.00	1.56	2.44
30°C	0.64	1.00	1.56
37°C	0.41	0.63	1.00

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures.

If control values are found outside the defined range, check the instrument, reagents and technique for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES¹

	25°C	30°C	37°C
Men, up to	80 U/L	130 U/L	195 U/L
Women, up to	70 U/L	110 U/L	170 U/L

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From detection limit of 2 U/L to linearity limit of 2400 U/L. If the results obtained were greater than linearity limit, dilute the sample 1/10 with NaCl 9 g/L and multiply the result by 10.

Precision:

Mean (U/L)	Intra-assay		Inter-assay	
	93	321	94	320
CV (%)	1.2	1.27	3.6	2.68

Sensitivity: 2 U/L.

Accuracy: Results obtained using our reagents (y) did not show systematic differences when compared with other commercial reagents (x).

The results obtained were the following:

Correlation coefficient (r): 0.986

Regression equation: y= 0.997x + 5.765

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

No interferences were observed with glucose until 7 g/L, bilirubin 400 µmol/L hemoglobin until 2.5 g/L and triglycerides 5g/L. A list of drugs and other interfering substances with CK determination has been reported by Young et. al^{3,4}.

NOTES

We have instruction sheets for several automatic analyzers. Instructions for many of them are available on request.

BIBLIOGRAPHY

1. Abbot B et al. Creatinine kinase. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984: 1112-1116.
2. Gerhardt W et al. Creatine kinase B-Subunit activity in serum after immunohinhibition of M-Subunit activity. Clin Chem 1979;(25/7): 1274-1280.
3. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
5. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
6. Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.
7. Mathieu M. et coll. Recommendation pour la mesure de la concentration catalytique de la créatinine kinase dans la sérum humain. Ann. Biol. Clin.,40, (1482), 87.